

EasySep™ Human Pan-Extracellular Vesicle Positive Selection Kit

For processing 20 mL of biofluid

Catalog #17891

Positive Selection

Document #1000005292 | Version 03



Scientists Helping Scientists™ | WWW.STEMCELL.COM

TOLL FREE PHONE 1 800 667 0322 • PHONE +1 604 877 0713

INFO@STEMCELL.COM • TECHSUPPORT@STEMCELL.COM

FOR GLOBAL CONTACT DETAILS VISIT OUR WEBSITE

Description

Isolate human extracellular vesicles (EVs) from plasma, serum, cell culture conditioned medium, and urine by immunomagnetic positive selection.

- Fast and easy-to-use
- No columns required

This kit targets EVs for positive selection with antibodies recognizing the specific tetraspanin markers CD9, CD63, and CD81. Desired EVs are labeled with antibodies and magnetic particles, and separated without columns using an EasySep™ magnet. Unwanted biofluid components are simply poured off, while desired EVs remain in the tube. The final isolated fraction contains highly purified EVs that are immediately available for downstream applications such as DNA/RNA extraction, western blot, or mass spectrometry.

Following positive selection, **particle release is not recommended**, and EVs remain bound to antibody complexes following particle release. Antibody complexes and particles bound to EVs may interfere with Billiant Violet™ antibody conjugates, polyethylene glycol-modified proteins, or other chemically related ligands.

Component Descriptions

COMPONENT NAME	COMPONENT #	QUANTITY	STORAGE	SHELF LIFE	FORMAT
EasySep™ Human Pan-Extracellular Vesicle Positive Selection Cocktail	17891C	1 x 1 mL	Store at 2 - 8°C. Do not freeze.	Stable until expiry date (EXP) on label.	A combination of monoclonal antibodies in PBS.
EasySep™ Releasable RapidSpheres™ 50201	50201	2 x 1 mL	Store at 2 - 8°C. Do not freeze.	Stable until expiry date (EXP) on label.	A suspension of magnetic particles in water.

PBS - phosphate-buffered saline

Components may be shipped at room temperature (15 - 25°C) but should be stored as indicated above.

Sample Preparation

For available fresh and frozen samples, see www.stemcell.com/primarycells.

PLASMA (FROM WHOLE BLOOD)

1. Centrifuge whole blood at 2000 x g for 10 minutes. Remove the plasma layer and transfer to a conical tube (e.g. Catalog #38009/38010).
2. Centrifuge the plasma layer (from step 1) at 2000 x g for 10 minutes. Remove the plasma supernatant and transfer to a new conical tube.
3. Centrifuge plasma supernatant at 10,000 x g for 30 minutes to remove cellular debris and large vesicles. Remove plasma supernatant and transfer to the required tube or plate (see Tables 1 - 2).

OPTIONAL: If desired, plasma can be filtered using a 0.2 µm filter prior to isolation of EVs.

For other biofluids (except urine), follow protocol for conditioned medium (steps 2 - 3) to remove cells and large vesicles.

CONDITIONED MEDIUM

1. Harvest conditioned medium from cell culture and transfer to a conical tube (e.g. Catalog #38009/38010).
2. Centrifuge the conditioned medium at 2000 x g for 10 minutes. Remove the supernatant and transfer to a new conical tube.
3. Centrifuge supernatant at 10,000 x g for 30 minutes. Remove the supernatant and transfer to the required tube or plate (see Tables 1 - 2).
4. OPTIONAL: If starting with diluted samples, concentrate using a 30K or 100K centrifugal filter tube (e.g. PALL Catalog #MAP030C36/MAP100C36).
5. OPTIONAL: If desired, supernatant can be filtered using a 0.2 µm filter prior to isolation of EVs.

URINE

If using frozen urine, thaw fully before processing the sample.

1. Vortex urine to obtain a homogenous suspension.
2. Centrifuge urine at 1000 x g for 10 minutes at room temperature (15 - 25°C). Remove supernatant and transfer to a new tube.

NOTE: For fresh urine, we recommend adding protease inhibitors to prevent protein degradation.

NOTE: If not used immediately, freeze urine at -20°C for long-term storage.

3. OPTIONAL (RECOMMENDED): Pre-clear sample by centrifuging supernatant at 10,000 x g for 30 minutes at room temperature.

NOTE: This step will reduce THP contamination, but may lower final EV recovery.

4. For samples ≤ 2 mL, remove supernatant and transfer to the required tube or plate (see Tables 1 - 2),

OR

For samples $> 2 - 20$ mL, transfer urine to a 100K centrifugal filter tube to concentrate the sample. Centrifuge at 1000 x g for 30 minutes at room temperature. Collect retained volume above filter membrane and transfer to the required tube (see Table 1). Top up to 1 mL with recommended medium.



Recommended Medium

D-PBS (Without Ca^{++} and Mg^{++} ; Catalog #37350)

Directions for Use – Manual EasySep™ Protocols

See pages 1 and 2 for Sample Preparation and Recommended Medium. Refer to Tables 1 and 2 for detailed instructions regarding the EasySep™ procedure for each magnet.

Table 1. EasySep™ Human Pan-Extracellular Vesicle Positive Selection Kit Protocol

		EASYSEP™ MAGNETS	
STEP	INSTRUCTIONS	 EasySep™ (Catalog #18000)	“The Big Easy” (Catalog #18001) 
1	Add sample to required tube.	≤ 0.5 - 2 mL	1 - 8 mL
	Required tube.	5 mL (12 x 75 mm) polystyrene round-bottom tube (e.g. Catalog #38007)	14 mL (17 x 95 mm) polystyrene round-bottom tube (e.g. Catalog #38008)
2	Add Selection Cocktail to sample. NOTE: Do not vortex cocktail.	50 µL/mL for samples ≥ 0.5 mL NOTE: For samples < 0.5 mL, add 25 µL of cocktail.	50 µL/mL of sample
	Mix and incubate.	RT for 10 minutes OR RT for 30 minutes for concentrated urine samples	RT for 10 minutes OR RT for 30 minutes for concentrated urine samples
3	Vortex Releasable RapidSpheres™. NOTE: Particles should appear evenly dispersed.	30 seconds	30 seconds
4	Add Releasable RapidSpheres™ to sample.	100 µL/mL for samples ≥ 0.5 mL NOTE: For samples < 0.5 mL, add 50 µL of particles.	100 µL/mL of sample
	Mix and incubate.	RT for 10 minutes OR RT for 30 minutes for concentrated urine samples	RT for 10 minutes OR RT for 30 minutes for concentrated urine samples
5	Add recommended medium to top up the sample to the indicated volume. Mix by gently pipetting up and down 2 - 3 times.	Top up to 2.5 mL	<ul style="list-style-type: none"> • Top up to 5 mL for samples < 4 mL • Top up to 10 mL for samples ≥ 4 mL
	Place the tube (without lid) into the magnet and incubate.	RT for 5 minutes	RT for 5 minutes
6	Pick up the magnet, and in one continuous motion invert the magnet and tube,* pouring off the supernatant. NOTE: Do not remove the tube from the magnet between separations.	Discard supernatant	Discard supernatant
7	Add recommended medium to top up the sample to the indicated volume. Mix by gently pipetting up and down 2 - 3 times.	Top up to 2.5 mL	<ul style="list-style-type: none"> • Top up to 5 mL for samples < 4 mL • Top up to 10 mL for samples ≥ 4 mL
	Incubate.	RT for 1 minute	RT for 1 minute
8	Pick up the magnet, and in one continuous motion invert the magnet and tube,* pouring off the supernatant. NOTE: Do not remove the tube from the magnet between separations.	Discard supernatant NOTE: If starting sample is conditioned medium, skip to step 10.	Discard supernatant NOTE: If starting sample is conditioned medium, skip to step 10.
9	Repeat steps as indicated.	Steps 7 and 8, two more times (total of 1 x 5-minute and 3 x 1-minute separations)	Steps 7 and 8, two more times (total of 1 x 5-minute and 3 x 1-minute separations)
10	Remove the tube from the magnet. Resuspend EVs in desired medium. Be sure to collect the EVs from the sides of the tube.	Isolated EVs are ready for use	Isolated EVs are ready for use

RT - room temperature (15 - 25°C)

* Leave the magnet and tube inverted for 2 - 3 seconds, then return upright. Do not shake or blot off any drops that may remain hanging from the mouth of the tube.

Table 2. EasySep™ Human Pan-Extracellular Vesicle Positive Selection Kit Protocol

STEP	INSTRUCTIONS	EasyPlate™ (Catalog #18102)
1	Add sample to required plate.	0.2 mL
2	Add Selection Cocktail to sample. NOTE: Do not vortex cocktail.	50 µL/mL of sample NOTE: For samples < 0.2 mL, add 10 µL of cocktail.
	Mix and incubate.	RT for 10 minutes OR RT for 30 minutes for concentrated urine samples
3	Vortex RapidSpheres™. NOTE: Particles should appear evenly dispersed.	30 seconds
4	Add Releasable RapidSpheres™ to sample.	100 µL/mL of sample NOTE: For samples < 0.2 mL, add 20 µL of RapidSpheres™.
	Mix and incubate.	RT for 10 minutes OR RT for 30 minutes for concentrated urine samples
5	Add recommended medium to top up the sample to the indicated volume. Mix by gently pipetting up and down 2 - 3 times.	Top up to 0.2 mL
	Place the plate (without lid) into the magnet and incubate.	RT for 5 minutes
6	Carefully pipette (do not pour) off the supernatant. NOTE: Do not remove the plate from the magnet between separations.	Discard supernatant
7	Add recommended medium to top up the sample to the indicated volume. Mix by gently pipetting up and down 2 - 3 times.	Top up to 0.2 mL
	Incubate.	RT for 1 minute
8	Carefully pipette (do not pour) off the supernatant. NOTE: Do not remove the plate from the magnet between separations.	Discard supernatant NOTE: If starting sample is conditioned medium, skip to step 10.
9	Repeat steps as indicated.	Steps 7 and 8, two more times (total of 1 x 5-minute separation and 3 x 1-minute separations)
10	Remove the plate from the magnet. Resuspend EVs in desired medium.	Isolated EVs are ready for use

RT - room temperature (15 - 25°C)

Notes and Tips

For assessment of CD9, CD63, and CD81 tetraspanin markers by western blot immunostaining, use Extracellular Vesicle Human CD9/CD63/CD81 Antibody Panel (Catalog #100-0211) or the following antibody clones:

- Anti-Human CD9 Antibody, Clone HI9A (Catalog #100-0138), and
- Anti-Human CD63 Antibody, Clone H5C6 (Catalog #100-0139), and
- Anti-Human CD81 (TAPA-1) Antibody, Clone 5A6 (Catalog #100-0209)

For more information, refer to the web protocol: How to Characterize Extracellular Vesicles by Western Blotting, available at www.stemcell.com.

BIOFLUID VARIABILITY


Types and levels of tetraspanin expression on EVs within and between biofluid samples can be variable. This may affect isolation yields and tetraspanin data obtained in subsequent analyses.

OPTIMIZING RECOVERY

To improve EV recovery, use one of the following methods:

- Increase Selection Cocktail incubation time to 30 minutes, and/or
- Double the volumes of Selection Cocktail and RapidSpheres™
NOTE: The volumes of the two components must be increased at the same time.
- For samples < 2 mL, use of EasySep™ Magnet (Catalog #18000) is recommended.

Data



The western blot analyses in the above examples show positive isolation of EVs with CD9, CD63, and CD81 tetraspanin markers from (A) human plasma and (B) conditioned medium.

PRODUCTS ARE FOR RESEARCH USE ONLY AND NOT INTENDED FOR HUMAN OR ANIMAL DIAGNOSTIC OR THERAPEUTIC USES UNLESS OTHERWISE STATED. FOR ADDITIONAL INFORMATION ON QUALITY AT STEMCELL, REFER TO WWW.STEMCELL.COM/COMPLIANCE.

Copyright © 2022 by STEMCELL Technologies Inc. All rights reserved including graphics and images. STEMCELL Technologies & Design, STEMCELL Shield Design, Scientists Helping Scientists, EasySep, EasyPlate, and RapidSpheres are trademarks of STEMCELL Technologies Canada Inc. All other trademarks are the property of their respective holders. While STEMCELL has made all reasonable efforts to ensure that the information provided by STEMCELL and its suppliers is correct, it makes no warranties or representations as to the accuracy or completeness of such information.