



"The Big Easy" EasySep™ Magnet

Magnet for column-free immunomagnetic separation

1 x 10⁹ cells

Catalog # 18001



Scientists Helping Scientists™ | WWW.STEMCELL.COM

TOLL FREE PHONE 1 800 667 0322 • PHONE +1 604 877 0713

INFO@STEMCELL.COM • TECHSUPPORT@STEMCELL.COM

FOR GLOBAL CONTACT DETAILS VISIT OUR WEBSITE

Product Description

"The Big Easy" EasySep™ Magnet is an anodized aluminum body housing rare earth magnetic elements.

"The Big Easy" EasySep™ Magnet is designed for cell separation procedures using EasySep™ reagents when processing a large number of cells (up to 1 x 10⁹ cells), or when using RoboSep™. "The Big Easy" EasySep™ Magnet generates a high-gradient magnetic field in the interior cavity that is strong enough to separate cells labeled with magnetic EasySep™ particles without the use of columns. "The Big Easy" EasySep™ Magnet is designed to hold a standard 14 mL (17 x 95 mm) polystyrene round-bottom tube (e.g. Catalog #38008).

Properties

Storage: Store at 15 - 25°C.

Shelf Life: Not applicable

STRONG MAGNETIC FIELD: Keep away from pacemakers, magnets, computer disks, watches, and other objects that respond to magnetic fields.

Handling / Directions For Use

"The Big Easy" EasySep™ Magnet is only for use with certain EasySep™ kits. For more information, refer to the appropriate EasySep™ Product Information Sheet or contact us at techsupport@stemcell.com.

The EasySep™ Magnet is designed to hold a 14 mL (17 x 95 mm) polystyrene round-bottom tube (e.g. Catalog #38008).

The size of the O-ring was selected for polystyrene tubes. Polystyrene tubes were chosen because recovery is typically better than with polypropylene tubes. Some customers may prefer to use polypropylene tubes because monocytes may adhere to polystyrene. If polypropylene tubes are used, we advise that the diameter of the opening be reduced by inserting a small piece of rubber band between the O-ring and the housing.

CARE OF MAGNET: After completing a cell separation, wipe the magnet clean with distilled water and then with alcohol to dry the magnet. To disinfect the magnet, wipe with 10% diluted household bleach and then wipe off with a water-dampened cloth to avoid corrosion of metal parts. Do not submerge the magnet in liquid.



STEMCELL TECHNOLOGIES INC.'S QUALITY MANAGEMENT SYSTEM IS CERTIFIED TO ISO 13485. PRODUCTS ARE FOR RESEARCH USE ONLY AND NOT INTENDED FOR HUMAN OR ANIMAL DIAGNOSTIC OR THERAPEUTIC USES UNLESS OTHERWISE STATED.

Copyright © 2017 by STEMCELL Technologies Inc. All rights reserved including graphics and images. STEMCELL Technologies & Design, STEMCELL Shield Design, Scientists Helping Scientists, RoboSep, and EasySep are trademarks of STEMCELL Technologies Canada Inc. All other trademarks are the property of their respective holders. While STEMCELL has made all reasonable efforts to ensure that the information provided by STEMCELL and its suppliers is correct, it makes no warranties or representations as to the accuracy or completeness of such information.