

Trypan Blue

Reagent for counting viable mammalian cells

Catalog # 07050 100 mL



Scientists Helping Scientists™ | WWW.STEMCELL.COM

TOLL FREE PHONE 1 800 667 0322 • PHONE +1 604 877 0713

INFO@STEMCELL.COM • TECHSUPPORT@STEMCELL.COM

FOR GLOBAL CONTACT DETAILS VISIT OUR WEBSITE

Product Description

Trypan Blue is recommended for counting viable mammalian cells. Viable cell counts should be performed when a decrease in cell viability may be expected, for example, when working with cryopreserved cells or cells manipulated ex vivo.

Properties

- Storage:** Store at 15 - 25°C.
Shelf Life: Stable until expiry date (EXP) on label.
Contains:
- Trypan blue (0.4%)
 - Phosphate-buffered saline

Please refer to the Safety Data Sheet (SDS) for hazard information.

Handling / Directions For Use

1. Dilute cells 1:1 in Trypan Blue.
NOTE: If the cell count appears high, the cells may first be diluted with a balanced salt solution (e.g. D-PBS [Without Ca⁺⁺ and Mg⁺⁺], Catalog #37350) before Trypan Blue is added.
2. Allow the resulting solution to sit for 5 - 15 minutes. Only non-viable cells will stain with the Trypan Blue dye; viable cells will remain unstained.
NOTE: If cells are incubated for greater than 15 minutes in Trypan Blue, toxicity effects may occur and the viable cell count will be inaccurate.
3. Prepare a hemocytometer by first cleaning the chamber surface with alcohol. Wipe dry.
4. Position the coverslip over the chambers. Carefully transfer sufficient volume of the Trypan Blue/cell solution to each chamber using a capillary tube or pipetman. Do not over- or underfill.
5. Count the cells in one chamber. Keep a separate count of viable (unstained) and non-viable (blue) cells. Count all cells in each 1 mm square of each chamber. If cells are on the border outlining each square, count only the cells on the top and left border of the square.
NOTE: Each square has a total volume of 0.1 mm³ (or 10⁻⁴ cm³, which is approximately equivalent to 10⁻⁴ mL).
6. Determine the cell count (cells per mL) as follows:
average cell count per square x dilution factor x 10⁴ = cell count per mL
7. Determine the cell viability (%) as follows:
(cell count [viable]) / (total cell count [viable + non-viable]) = cell viability (%)

STEMCELL TECHNOLOGIES INC.'S QUALITY MANAGEMENT SYSTEM IS CERTIFIED TO ISO 13485. PRODUCTS ARE FOR RESEARCH USE ONLY AND NOT INTENDED FOR HUMAN OR ANIMAL DIAGNOSTIC OR THERAPEUTIC USES UNLESS OTHERWISE STATED.

Copyright © 2016 by STEMCELL Technologies Inc. All rights reserved including graphics and images. STEMCELL Technologies & Design, STEMCELL Shield Design, and Scientists Helping Scientists are trademarks of STEMCELL Technologies Canada Inc. All other trademarks are the property of their respective holders. While STEMCELL has made all reasonable efforts to ensure that the information provided by STEMCELL and its suppliers is correct, it makes no warranties or representations as to the accuracy or completeness of such information.