

HemaTox™ Erythroid Kit

Serum-free medium and supplement for the measurement of drug toxicity on erythroid progenitor cells

Catalog #09701

1 Kit



Scientists Helping Scientists™ | WWW.STEMCELL.COM

TOLL FREE PHONE 1 800 667 0322 • PHONE +1 604 877 0713

INFO@STEMCELL.COM • TECHSUPPORT@STEMCELL.COM

FOR GLOBAL CONTACT DETAILS VISIT OUR WEBSITE

Product Description

HemaTox™ Erythroid Kit is intended for the measurement of drug toxicity on the proliferation and erythroid-specific differentiation of human hematopoietic stem and progenitor cells (HSPCs) in a liquid culture-based 96-well plate format.

HemaTox™ Erythroid Kit includes a specialized serum-free culture medium and 100X supplement. Complete HemaTox™ Erythroid Medium (HemaTox™ Erythroid Medium + HemaTox™ Erythroid 100X Supplement) promotes the proliferation of human CD34+ HSPCs and their differentiation into erythroid cells during a 7-day culture period. After culture, the cells can be counted and assessed for expression of erythroid markers such as the transferrin receptor (CD71) and Glycophorin A (GlyA, CD235a) using flow cytometry or other methods.

HemaTox™ Erythroid Kit may be used on its own or in combination with HemaTox™ Myeloid Kit (Catalog #09704) or HemaTox™ Megakaryocyte Kit (Catalog #09707) to assess lineage-specific drug toxicity in parallel.

Each kit contains sufficient medium and supplement for testing up to 160 different conditions (triplicate wells per condition, 200 µL per well) in 5 x 96-well plates.

Product Information

The following components are sold as a complete kit (Catalog #09701) and are not available for individual sale.

COMPONENT NAME	COMPONENT #	SIZE	STORAGE	SHELF LIFE
HemaTox™ Erythroid Medium*	09702	100 mL	Store at -20°C.	Stable until expiry date (EXP) on label.
HemaTox™ Erythroid 100X Supplement	09703	1 mL	Store at -20°C.	Stable until expiry date (EXP) on label.

*This product contains material derived from human plasma. Donors have been tested and found negative for HIV-1 and -2, hepatitis B, and hepatitis C prior to donation. However, this product should be considered potentially infectious and treated in accordance with universal handling precautions.

Materials Required But Not Included

PRODUCT NAME	CATALOG #
96-well flat-bottom plates	38022
245 mm x 245 mm Square Treated Tissue Culture Dishes	27140
35 mm Culture Dishes	27100
Solvent for test compound(s), e.g. dimethyl sulfoxide (DMSO)	--
10% Bovine Serum Albumin (BSA) in Iscove's MDM	09300
Iscove's Modified Dulbecco's Medium (IMDM)	36150
Trypan Blue	07050

Preparation of Reagents and Materials

Complete HemaTox™ Erythroid Medium

Use sterile techniques to prepare **complete** HemaTox™ Erythroid Medium (HemaTox™ Erythroid Medium + HemaTox™ Erythroid 100X Supplement). The following example is for preparing 20 mL of complete medium. If preparing other volumes, adjust accordingly.

NOTE: Prepare only the volume of **complete** HemaTox™ Erythroid Medium required for a single experiment. A 20 mL volume is sufficient for one 96-well plate at 200 µL/well.

1. Thaw HemaTox™ Erythroid Medium at 2 - 8°C overnight. Mix well.
NOTE: Once thawed, use immediately or aliquot and store at -20°C until the expiry date (EXP) as indicated on the label. After thawing the aliquots, use immediately. Do not re-freeze.
2. Thaw HemaTox™ Erythroid 100X Supplement at room temperature (15 - 25°C) immediately prior to use. Mix well.
NOTE: Once thawed, use immediately or aliquot and store at -20°C until the expiry date (EXP) as indicated on the label. After thawing the aliquots, use immediately. Do not re-freeze.
3. Add 200 µL of HemaTox™ Erythroid 100X Supplement to 19.8 mL of HemaTox™ Erythroid Medium. Mix thoroughly.

Test Compound Solutions

1. Dissolve or dilute the test compound in an appropriate solvent, preferably to at least 1000X the concentration at which it will be tested in culture. This is the test compound stock solution. Different dilutions may need to be prepared depending on the solubility of the test compound and the required concentration range.
NOTE: A stock solution concentration $\geq 1000X$ will ensure that the final solvent concentration in the culture will be $\leq 0.1\%$. If DMSO is the solvent, a DMSO concentration of $\leq 0.1\%$ will not affect cell growth.
2. Prepare a 2X test compound solution by diluting the test compound stock solution (prepared in step 1) in **complete** HemaTox™ Erythroid Medium.
NOTE: Prepare sufficient volume of 2X test compound solution for replicate wells at 100 µL/well. Three replicate wells are recommended for each test compound.
3. Prepare a solvent control by diluting solvent in **complete** HemaTox™ Erythroid Medium to the same concentration as the solvent in the 2X test compound solution.

Isolation of Fresh CD34+ Cells from Cord Blood or Bone Marrow

Drug toxicity assays should be performed using purified human CD34+ cells isolated from cord blood (CB) or bone marrow (BM). The use of unfractionated or minimally processed CB or BM is not recommended.

CD34+ cells can be isolated from whole CB using EasySep™ Human Cord Blood CD34 Positive Selection Kit II (Catalog #17896) or from BM mononuclear cells using EasySep™ Human CD34 Positive Selection Kit (Catalog #18056). Freshly isolated CD34+ cells can be used immediately or aliquots may be cryopreserved for later use.

Directions for Use

Please read the entire protocol before proceeding.

A. THAWING, WASHING, AND DILUTING CD34+ CELLS

1. Thaw the vial of cells quickly (within 2 minutes) in a 37°C water bath by swirling gently.
2. When the cells are almost completely thawed, wipe the outside of the vial with 70% ethanol or isopropanol.
3. Transfer cells to a 15 mL or 50 mL tube.
4. Slowly (dropwise) add wash medium (e.g. 1% BSA in Iscove's MDM) to the thawed cells to a final volume of 10 mL or 50 mL while gently swirling the tube (approximately 1 - 2 minutes). Invert the tube to mix.
5. Centrifuge the cell suspension at $300 \times g$ for 10 minutes at room temperature (15 - 25°C).
6. Carefully remove the supernatant, taking care not to dislodge the cell pellet. Do not pour off.
7. Resuspend the cell pellet by gently flicking the tube. Add a known volume of either IMDM or **complete** HemaTox™ Erythroid Medium.
8. Perform a viable cell count using Trypan Blue and a hemocytometer, or use an automated cell counter.

NOTE: Methods to assay viable cells (e.g. dye exclusion) should be used for cell preparations where a decrease in cell viability may be expected (e.g. cryopreserved cells).

9. Dilute the cells in **complete** HemaTox™ Erythroid Medium to a concentration of 1000 viable CD34+ cells/100 µL (10,000 cells/mL).
NOTE: Prepare at least 10 mL of cell suspension for each 96-well plate to ensure that there is a sufficient volume to seed the required number of wells for the experiment (100 µL/well).

B. PLATING

- Mix the cell suspension (prepared in section A) immediately before use. Add 100 µL of the cell suspension to each well of the 96-well plate.
- Add 100 µL of the appropriate 2X test compound solution or solvent control to each well.
NOTE: Three replicate wells are recommended for each test compound.
- Place each 96-well plate in a 245 mm x 245 mm Square Treated Tissue Culture Dish. Within this outer tissue culture dish, surround the 96-well plate with 4 x 35 mm Culture Dishes containing ~3 mL of sterile water.
- Incubate at 37°C in 5% CO₂ and > 95% humidity for 7 days.

C. CULTURE ANALYSIS

Choose an appropriate analysis method for determining the effect of test compounds on HSPC proliferation and erythroid differentiation.

We recommend labeling cells with antibodies for cell surface markers that characterize erythroid cell populations, then counting cells that express the markers using a flow cytometer with absolute cell counting capability. Refer to Table 1 for recommended antibodies.

Table 1. Recommended Antibodies for Labeling Erythroid Cells

ANTIBODY*	CATALOG #	LABELED CELL POPULATIONS
Anti-Human CD71 (Transferrin Receptor) Antibody, Clone OKT9, PE	60106PE	<ul style="list-style-type: none"> Erythroid progenitor cells/ pro-erythroblasts (CD71+GlyA-) Erythroblasts (CD71+GlyA+) Normoblasts/reticulocytes (CD71^{lo}GlyA+)
Anti-Human CD235a (Glycophorin A) Antibody, Clone 2B7, FITC	60152FI	

*Recommended fluorochromes are shown; however, users should select the fluorochrome that is appropriate for their instrument and analysis.

Alternative analysis methods (e.g. automated cell counting, imaging cytometry, or plate reader-based methods) may also be used to quantify the response and obtain estimates for the 50% and 90% inhibitory concentrations (IC₅₀ and IC₉₀) for each test compound in the assay. Optimization and validation may be required for each analysis method.

Related Products

For related products, including specialized culture and storage media, supplements, antibodies, cytokines, and small molecules, visit www.stemcell.com/HSPCworkflow or contact us at techsupport@stemcell.com. For available fresh and cryopreserved peripheral blood, cord blood, and bone marrow products, visit www.stemcell.com/primarycells.

STEMCELL TECHNOLOGIES INC.'S QUALITY MANAGEMENT SYSTEM IS CERTIFIED TO ISO 13485. PRODUCTS ARE FOR RESEARCH USE ONLY AND NOT INTENDED FOR HUMAN OR ANIMAL DIAGNOSTIC OR THERAPEUTIC USES UNLESS OTHERWISE STATED.

Copyright © 2018 by STEMCELL Technologies Inc. All rights reserved including graphics and images. STEMCELL Technologies & Design, STEMCELL Shield Design, Scientists Helping Scientists, EasySep, and HemaTox are trademarks of STEMCELL Technologies Canada Inc. All other trademarks are the property of their respective holders. While STEMCELL has made all reasonable efforts to ensure that the information provided by STEMCELL and its suppliers is correct, it makes no warranties or representations as to the accuracy or completeness of such information.